


**ECG MACHINE THREE CHANNEL (APPROVED PVMS)  
FOR THE FINANCIAL YEAR 2018-19  
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD.**

TECHNICAL SPECIFICATIONS	Qty
<p>Three channel ECG on at least 3 inches LCD display.            Display of Three channel ECG simultaneously.            Automatic Operation            Variable gain: 1/2, 1, 2 cm/mV            Thermal recorder for printing out of Three channels simultaneously            Interpretation Software.            Recording Trace speed: 10, 25 and 50 mm/sec            Muscle artifact and AC (50Hz) interference filters            Defibrillator protection            Built in AC Supply and battery operation with 30min backup            Built-in AC interference, Noise filter and Baseline connection.            Capability to interface with LAN/WLAN for data transfer            Paper Rolls, 50</p>	10 Nos.
<p><b>Accessories:</b>            Complete with standard accessories, including separate patient cables for Adult, Pediatric use with re-usable electrodes</p>	
<p>Optional (If any):            Mobile Cart (Local/Imported)</p>	

  
 \_\_\_\_\_  
**Medical Superintendent**  
 Faisalabad Institute of Cardiology,  
 Faisalabad.

