

MINUTES OF THE TECHNICAL ADVISORY COMMITTEE MEETING HELD ON 20.03.2017
AT 11:00 AM IN THE CONFERENCE ROOM, F.I.C FAISALABAD FOR PURCHASE OF
BIOMEDICAL ACCESSORIES ITEMS, F.I.C FAISALABAD

FOLLOWING ATTENDED

1.	Head of Cardiology, F.I.C. Faisalabad	Chairman
2.	A.M.S (Purchase), F.I.C. Faisalabad	Member
3.	D.M.S (Purchase), F.I.C. Faisalabad	Member
4.	Bio Medical Engineer, F.I.C. Faisalabad	Member
5.	Pharmacist, F.I.C. Faisalabad	Member
6.	Dr. Irfan, Medical Officer F.I.C. Faisalabad	Copt. Member

The committee was intimated that tender was invited through national press i-e "Daily JANG" 24.02.2017 and on PPRA dated 25-02-2017 with due date of receiving tender as 16-03-2017. The tender was opened on the same day in the presence of bidders/ their authorized representative. Following firms participated in the tender process with their technical offers. The decisions of TAC are under.

Award #	Name of Item	Brand with objection	Contractor/ Firm
1	Rechargeable Battery Ventilator Servo-i	Italy TNA Low quality	M/s Veroline Engineers
		Compatible Explore Synergy Inc. Hong Kong TNA Sample not provided	M/s Novatek Pakistan
		Battery will be refurbished using imported cell from china, battery casing will be provided from end user. warranty 1 year TNA Sample not provided	M/s Delta Plus Pvt. Ltd.
		Maquet Sweden T.A	M/s Digionics Pvt. Ltd.
		USA TNA Sample not provided	M/s Pearls
2	Air Suction Venturi complete alongwith Jar and fitting USA/EU/Japan	UK TNA Low quality	M/s Veroline Engineers
		GCE UK TNA Low quality	M/s Novatek Pakistan
		ITC-Spain (Europe) TNA Sample not provided	M/s Delta Plus Pvt. Ltd.
		EASY AIR 1000 with MAK 1000 jar Flowmeter-Italy T.A	M/s Total Technologies Pvt. Ltd.
		ITC- Spain TNA Sample not provided	M/s Pearls
		Technologie Medicale- France TNA Low quality	M/s Med Impex
3	Maintenance Kit of Servo-I Ventilator 5000 Hrs Maquet	Maquet Sweden T.A	M/s Digionics Pvt. Ltd.

[Handwritten signatures and initials in blue ink]

