

**MINUTES OF THE MEETING REGARDING RE-EVALUATION OF CARDIAC SURGICAL DISPOSABLE ITEMS
NO. 3,5,6,7,19,20,21,22,23,24,26,62,68,76,78,88,91,93,146 AND 147 HELD ON 11.11.2020 IN THE
CONFERENCE ROOM, FIC, FAISALABAD FOR THE F.Y 2020-2021.**

Following members attended the meeting.

- 1 **Dr. Hamid Saeed**, Head of Cardiology, F.I.C. Faisalabad.
- 2 **Dr. Muhammad Sajid**, AMS (Purchase/Stores), F.I.C. Faisalabad.
- 3 **Dr. Husnain Zafar**, Assistant Professor of Cardiac Surgery, FIC, Faisalabad
- 4 **Dr. Zaigam Rasool**, Assistant Professor of Cardiac Surgery (Peads), FIC, Faisalabad.
- 5 **Dr. Asghar Ali**, Anesthetist, FIC, Faisalabad.
- 6 **Miss. Iqra Liaqat**, Pharmacist, F.I.C. Faisalabad.
- 7 **Mr. Najam Ul Hassan**, Electronic Equipment Care Taker, F.I.C. Faisalabad.

Agenda:

The meeting was called on the recommendation of Grievance committee for the re-evaluation of following products.

1. M/s Imtiaz Brothers				
Item #	Item Name	Brand Name	Manufactured By / Imported By	Remarks
78	OXYGEN MASK ALL SIZES	Max	Anhui Ganor Medical Device Co, Ltd. China	T.A
2. M/s Anax Associates				
21	CARDIO VASCULAR DRAPE SET	Medpro Cardio Vascular Drape Set	Medpro, China	T.N.A Poor quality
3.M/s Husnain Pharma				
3	ADHESIVE DRESSING ROLL 4" (10CM x 10M/9.14M)	Hospital & Homecare (10CMX10M)	China	T.N.A Poor Stickiness
19	CANNULA NASAL FOR OXYGEN DELIVERY -ADULT	Flexicare	Flexicare Medical UK	T.N.A 1.Punjab Govt. experience not attached 2. Could not maintain desired flow rate
76	NEBULIZER KITS ALL SIZES	Maxineb	Flexicare Medical Limited UK	T.N.A 1.Punjab Govt. experience not attached 2. Could not maintain desired flow rate



