## MINUTES OF THE TECHNICAL ADVISORY COMMITTEE MEETING HELD ON 22.10.2018 AT 11:00 AM IN THE CONFERENCE ROOM, F.I.C FAISALABAD FOR PURCHASE OF ECG MACHINE FOR F.I.C FAISALABAD

## **FOLLOWING ATTENDED**

Head of Cardiology, F.I.C. Faisalabad

Chairman

2. A.M.S (Purchase), F.I.C. Faisalabad Member

3. D.M.S (Purchase), F.I.C. Faisalabad Member

4.

Electronic Equipment Care Taker, F.I.C. Faisalabad

Member

5. Pharmacist, F.I.C. Faisalabad

Member

The Committee was intimated that tender was invited through national press i-e "Daily Jang" 06.09.2018 "Daily The Nation 06.09.2018" and on PPRA website dated 08.09.2018 with due date of receiving tender as 26.09.2018. The tender was opened on the same day in the presence of bidders or their authorized representative. Following firms participated in the tender process

1. M/s Delta Plus

with their technical offers.

- 2. M/s Hospicare Systems
- 3. M/s Shirazi Trading
- 4. M/s Orient Medical
- 5. M/s KASBN International
- M/s Delta Plus & M/s Hospicare Systems were declared responsive by the technical advisory committee.
- M/s Orient Medical technically rejected because firm has No product legibility and No LAN/WLAN.
- M/s Shirazi Trading technically rejected because firm has No product legibility, No LAN/WLAN, No builtin AC supply.
- > M/s KASBN International technically rejected because Lumed is whole sale dealer instead of manufacturer as mentioned on their attached certification.
- Hence M/s Orient Medical, M/s Shirazi Trading and M/s KASBN International were declared Non Responsive

**Pharmacist** 

F.I.C. Faisalabad

**Electronic Equipment** 

Care Taker

F.I.C. Faisalabad

F.I.C. Faisalabad

A.M.S (Purchase),

F.I.C. Faisalabad

Head of Cardiology

F.I.C. Faisalabad