

**MINUTES OF GRIEVANCE COMMITTEE MEETING
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD**

A meeting of Grievance Committee for product of Lab Kits & Chemicals Items for F.Y. 2020-21 was held on 11.11.2020 at 10:00AM in the Office of the Medical Superintendent, Faisalabad Institute of Cardiology, Faisalabad.

Lab Kits & Chemicals Items

1. M/s DiaScient International

Item #	Item Name	Brand Name	TAC Observations	GC Observation	Remarks
48	Troponin I Devices	Dia Sign	Invalid Authority Letter Attached No Product Evaluation (Sample not Provided)	valid Authority Letter Attached No Product Evaluation due to non provision of sample	Rejected
49	HBsAg- Screening Method	Dia Sign	Invalid Authority Letter Attached No Product Evaluation (Sample not Provided)	valid Authority Letter Attached No Product Evaluation due to non provision of sample	Rejected
50	Anti HCV Ab- Screening method	Dia Sign	Invalid Authority Letter Attached No Product Evaluation (Sample not Provided)	valid Authority Letter Attached No Product Evaluation due to non provision of sample	Rejected
51	Typhidot igG+IgM Screening	Dia Sign	Invalid Authority Letter Attached No Product Evaluation (Sample not Provided)	valid Authority Letter Attached No Product Evaluation due to non provision of sample	Rejected
52	Urine Strips	Dia Sign/ Excel	Invalid Authority Letter Attached	Valid Authority Letter Attached	Accepted
68	Diatron Diluent 20 liters(5 part differential) With daily normal Controls, periodic maintenance, consumables	Dymind Biotech	Invalid Authority Letter Attached These reagents do not comply with analyzer	Valid Authority Letter Attached Reagents do not comply with analyzer	Rejected
69	Diatron Lyse 05 Liters(5 part differential) With daily normal Controls, periodic maintenance, consumables	Dymind Biotech	Invalid Authority Letter Attached These reagents do not comply with analyzer	Valid Authority Letter Attached Reagents do not comply with analyzer	Rejected

