

**MINUTES OF THE TECHNICAL EVALUATION COMMITTEE MEETING HELD ON 21-07-2020 AT 11:00 AM IN THE CONFERENCE ROOM OF, F.I.C, FAISALABAD FOR PRODUCT EVALUATION OF ANGIOGRAPHY / ANGIOPLASTY DISPOSABLE ITEMS FOR THE FINANCIAL YEAR 2020-2021.**

Following attended

- |     |   |              |
|-----|---|--------------|
| 01. | Head of Cardiology, F.I.C, Faisalabad.              | Chairman     |
| 02. | AMS (Purchase/Stores), F.I.C, Faisalabad.           | Member       |
| 03. | Dr. Rehan Riaz, F.I.C. Faisalabad.                  | Copt. Member |
| 04. | Dr. Abdul Razzaq Mughal, F.I.C. Faisalabad.         | Copt. Member |
| 05. | Pharmacist, F.I.C. Faisalabad.                      | Member       |
| 06. | Electronic Equipment Care Taker, F.I.C. Faisalabad. | Member       |

**Proceeding of Meeting:**

Meeting started with recitation from Holy Quran. The committee was intimated that tenders were invited through DGPR. Tender notice appeared in Daily The Nation dated 04 June 2020, Daily The Express dated 04 June 2020 and on PPRA website dated 06 June 2020 for the Purchase of Angiography / Angioplasty Disposable. The matter discussed and decisions made are given as under: List of items / brands which were technically approved / rejected are as under.

Sr. No.	Award #	Name of Item	Brand / Manufacturer	Contractor/ Firm
1.	1	ARTERIAL PUNCTURE NEEDLE (FEMORAL) 18G	T.A Intradyn Puncture Needle 18G B-Braun Pakistan	B-Braun Pakistan
2.	2	MULLEN TRANSPETAL CATHETER INTRODUCER SHEATH ✓	T.N.A Medical device registration certificate not attached Cook Medical USA / Europe	Verizon
3.	3	BROKEN BROUGH TRANSPETAL NEEDLE ✓	T.N.A Medical device registration certificate not attached Cook Medical USA / Europe	Verizon
4.	4	CLOT ASPIRATION CATHETER ✓	T.N.A 1.Invalid Authority letter attached 2.Free sale certificate not attached Thrombuster II Kaneka, Japan	Intek Corporation
			T.A Advance Export Galway, Ireland. Medtronic	Medtronic
5.	5	CORONARY STENT - DRUG ELUTING BUDGET STENTS FDA APPROVED (All diameters UPTO 38MM) According to DRAP Notification ✓	T.N.A 1.Invalid Drug Registration Certificate attached 2. Medical device registration not attached Promus Element Plus Boston Scientific USA	Ferozsons Laboratories Limited
			T.A Resolute Integrity Galway, Ireland Medtronic	Medtronic
6.	6	STANDARD TOP OF THE LINE, DRUG ELUTING STENTS 2nd Generation, FDA APPROVED (ALL LENGTH) According to DRAP Notification ✓	T.A Promus Premier Up to 38mm Boston Scientific USA	Ferozsons Laboratories Limited
			T.N.A The brand quoted by the firm falls in category of 3 <sup>rd</sup> generation top of the line stents according to the DRAP notification while the firm quoted this brand in the category of 2 <sup>nd</sup> generation	Digital Imaging Systems

*[Handwritten signatures and initials]*

			of stents Xience Xpedition Abbott vascular, USA/Europe	
			T.N.A The brand quoted by the firm falls in category of 3rd generation top of the line stents according to the DRAP notification while the firm quoted this brand in the category of 2nd generation of stents Resolute Onyx Galway, Ireland Medtronic	Medtronic
7.	7	DRUG ELUTING STENTS (ALL LENGTH & ALL SIZES) a) TOP OF THE LINE LATEST USA, FDA APPROVED 3rd Generation b) EUROPEAN CE MARK and Japanese FDA Approved According to DRAP Notification	T.N.A According to DRAP notification The brand Coroflex ISAR fall in the category of these stents while the company offer Coroflex ISAR Neo brand in this category Coroflex ISAR Neo Coronary Stent System B-Braun Pakistan	B-Braun
			T.A Xience Xpedition Abbott vascular, USA/Europe	Digital Imaging Systems
			T.A Cre8 CID, Itlay	ATCO Pharma International Pvt Ltd.
8.	8	COVERED STENT - ALL SIZES	T.N.A 1. Medical device registration certificate not attached 2. Technically accepted subject to provision of sample. Bentley Germany	Verizon
			T.A Infinity Cordis USA	Global Marketing Services
			T.N.A 1. Invalid drug registration certificate attached 2. Medical device registration certificate not attached Alvision Alvimedica, Turkey	ATCO Pharma International Pvt Ltd.
9.	9	DIAGNOSTIC CATHETER ALL SIZES & ALL TYPES	T.A Pointer Pendraccare, Netherland	Health Tec
			T.A Dexterity Diagnostic Catheter Medtronic USA	Medtronic
			T.A Angiodyn Diagnostic Catheter Pendraccare International/ B-Braun Pakistan	B-Braun
10.	10	DISTAL PROTECTION DEVICE (FILTER TYPE) ALL SIZES	T.N.A 1. Medical device registration not attached 2. Invalid Provisional certificate attached Filter wire EZ Boston Scientific USA	Ferozsons Laboratories Limited
			T.A Protégé Blue Medical, Netherland	Health Tec.
11.	11	DRUG ELUTING BALLOON ALL SIZES	T.N.A All lengths are not available In.Pact Falcon Galway, Ireland Medtronic	Medtronic
			T.A SeQuent Please Neo PTCA Drug Eluting Balloon B-Braun Pakistan	B-Braun

*[Handwritten signatures and initials]*

	12	FFR WIRE	T.N.A Medical device registration certificate not attached FFR WIRE Phillips/Volcano USA/Europe	Digital Imaging Systems
13.	13	GUIDE WIRE 150-160 CM STANDARD LENGTH 0.035" J TIP	T.A Accoat 150CM SP Medical Denmark	Health Tec
			T.N.A 1. Not according to specification (quoted 145cm) PTFE Guide Wire Medtronic USA	Medtronic
			T.N.A 1. Medical device registration certificate not attached 2. Punjab govt. experience not attached 3. No past experience with this institute 4. Free sale certificate not attached Multy Way Medical UK/CHN	Bestech Enterprises
			T.N.A Medical device registration certificate not attached Angiodyn B-Braun Pakistan	B-Braun
			T.N.A 1. Medical device registration certificate not attached 2. Not according to Specification (145CM quoted) Cook Medical USA / Europe	Verizon
			T.A Emerald Cordis USA	Global Marketing Services
14.	14	GUIDE WIRE EXCHANGE LENGTH 0.035" 260CM J TIP	T.A Accoat 260CM SP Medical Denmark	Health Tec
			T.A PTFE Guide Wire Medtronic USA	Medtronic
			T.N.A 1. Medical device registration certificate not attached 2. Punjab govt. experience not attached 3. No past experience with this institute 4. Free sale certificate not attached Multy Way Medical UK/CHN	Bestech Enterprises.
			T.N.A Medical device registration certificate not attached Angiodyn B-Braun Pakistan	B-Braun Pakistan
			T.N.A Medical device registration certificate not attached Cook Medical USA / Europe	Verizon
			T.A Emerald Cordis USA	Global Marketing Services
15.	15	GUIDE WIRE EXCHANGE LENGTH 0.025" 260CM J TIP	T.N.A Medical device registration certificate not attached Cook Medical USA / Europe	Verizon
16.	16	GUIDING CATHETER ALL SIZES & ALL TYPES	T.A Vista Brite Tip Cordis USA	Global Marketing Services.
			T.A Alviguide Blue Plus Alvimedica, Turkey	ATCO Pharma International Pvt Ltd

*[Handwritten signatures and initials]*

			<p>T.N.A In sufficient experience with this institute Premium Hydrophilic Coated Pendracare, Netherland</p>	Health Tec
			<p>T.N.A 1. Invalid free sale certificate attached Launcher Medtronic USA</p>	Medtronic
			<p>T.N.A 1. Stiff material 2. No experience with this institute Falcon Guiding Catheter Umbra Medical Products Inc. USA</p>	Ali Medical Store
			<p>T.A Priority Pack Abbott vascular, USA/Europe</p>	Digital Imaging Systems
			<p>T.N.A 1. Medical device registration certificate not attached 2. Not user friendly Perouse Medical Vygon France</p>	Cardiac Care
			<p>T.N.A 1. Not user friendly 2. Punjab Govt. experience not attached 3. Medical device registration certificate not attached 4. Invalid free sale certificate attached Blow Medas USA</p>	Superior Healthcare International
17.	17	INFLATION DEVICE WITH (Y-CONNECTOR) (INTRODUCER NEEDLE) (TORQUER)	<p>T.N.A 1. Authority letter of manufacturer M/s JingzhouYihai Technology attached while in technical offer manufacturer Y-H Medical mentioned 2. Medical device registration certificate not attached 3. Punjab govt. experience not attached 4. Free sale certificate not attached 5. Not user friendly 6. No past experience with this institute Y-H Medical China</p>	Bestech Enterprises
18.	18	IVUS CATHETER	<p>T.A IVUS CATHETER Phillips/Volcano USA/Europe</p>	Digital Imaging Systems
19.	19	LONG SHEATH ALL SIZES	<p>T.A Cook Medical USA / Europe</p>	Verizon
			<p>T.N.A Medical device registration certificate not attached Angiodyn- Manifold B-Braun Pakistan</p>	B-Braun Pakistan
			<p>T.A Namic Navilyst Medical, USA</p>	A & F Medical
21.	21	MICRO/SUPPORT HYBIRD CATHETER	<p>T.N.A 1. Medical Device Registration certificate not attached 2. Free sale certificate not attached FineCross Terumo, Japan</p>	Intek Corporation
22.	22	PERFUSION BALLOON - ALL SIZES	<p>T.N.A Medical device registration certificate not attached Cook Medical USA / Europe</p>	Verizon
23.	23	PERIPHERAL BALLOON (MONORAIL) - ALL SIZES	<p>T.A XXI Boston Scientific USA</p>	Ferozsons Laboratories Limited


24.	24	PERIPHERAL STENTS (SELF EXPANDABLE/ BALLOON EXPANADABLE) - ALL SIZES	<p>T.A Wall Stent Boston Scientific USA</p>	Ferozsons Laboratories Limited.
			<p>T.N.A 1. Medical device registration certificate not attached 2. Technically accepted subject to provision of sample Cook Medical USA / Europe</p>	Verizon
25.	25	PERMANENT PACEMAKER DUAL CHAMBER (ARTERIAL AND VENTRICLE LEAD +INTRODUCER SHEATS + BATTERY) DDDR	<p>T.A Sensia Medtronic Switzerland/Singapore</p>	Medtronic
26.	26	PERMANENT PACEMAKER SINGLE CHAMBER (INTRODUCER SHEATH + LEAD+BATTERY) VVIR	<p>T.A Sensia Medtronic Switzerland/Singapore</p>	Medtronic
27.	27	PRESSURE TUBING FOR HEMO DYNAMIC MONITORING	<p>T.N.A 1. Manufacturer name and country of origin not mentioned on technical offer 2. Punjab Govt. experience not attached 3. Sample not provided 4. Free sale certificate not attached 5. Authority letter not attached Venous Flux 120cmx150cm</p>	Superior Healthcare International
			<p>T.A Ultramed Ultra For Medical, Egypt</p>	Husnain Pharma
			<p>T.A Combidyn Pressure B-Braun Pakistan</p>	B-Braun Pakistan
			<p>T.A Pressure Line 800 PSI Bicakliar Dis Ticaret A.S Turkey</p>	Ali Medical Store
			<p>T.A Pro Active Italy</p>	Cardiac Care
28.	28	PTCA BALLOON MONORAIL NON-COMPLIANT-ALL SIZES	<p>T.A SeQuent Neo NC B-Braun Pakistan</p>	B-Braun Pakistan
			<p>T.N.A 1. Technically accepted subject to provision of Punjab govt. experience 2. Free Sale Certificate not attached Appoio Bros Med Co. Ltd Netherland</p>	Flowtronix Systems
			<p>T.A Hawk NC Balloon Umbra Medical Products Inc. USA</p>	Ali Medical Store
			<p>T.A NC Quantum Apex Boston Scientific USA</p>	Ferozsons Laboratories Limited
			<p>T.N.A 1. Discrepancy found in manufacturer name (The manufacturer name mentioned on DRAP Medical device registration was M/s KONEKA Corp. Japan while on technical offer it was mentioned by the firm M/s Cordis USA) 2. Authority letter for this product not attached 3. Free Sale Certificate not attached 4. Punjab govt. experience not attached Raiden 3 Cordis USA</p>	Global Marketing Services
			<p>T.A Force-NC Blue Medical, Netherland</p>	Health Tec
			<p>T.A Sprinter NC Balloon Mexico Medtronic</p>	Medtronic

~~Handwritten signature~~      ✓      ✓      ✓      ✓      ✓

			T.A SeQuent Neo B-Braun Pakistan	B-Braun Pakistan
			T.N.A 1. Technically accepted subject to provision of Punjab govt. experience 2. Free Sale Certificate not attached Artimes Bros Med Co. Ltd	Flowtronix Systems
			T.A Hawk SC Balloon Umbra Medical Products Inc. USA	Ali Medical Store
			T.A Maverick II Boston Scientific USA	Ferozsons Laboratories Limited
29.	29	PTCA BALLOON MONORAIL (SEMI COMPLIANT) - ALL SIZES	T.N.A 1. Discrepancy found in manufacturer name (The manufacturer name mentioned on DRAP Medical device registration was M/s KONEKA Corp. Japan while on technical offer it was mentioned by the firm M/s Cordis USA) 2. Authority letter for this product not attached 3. Free Sale Certificate not attached 4. Punjab govt. experience not attached Ikazukhi Zero Cordis USA	Global Marketing Services
			T.A Everest-Sc Blue Medical. Netherland	Health Tec
			T.A Sprinter Legend SC Mexico Medtronic	Medtronic
			T.N.A Punjab govt. experience not attached Fluydo CID. Italy	ATCO Pharma International Pvt Ltd
30.	30	PTCA BALOON (OTW) SEMI COMPLAINT ALL SIZES	T.A Emerge OTW Boston Scientific USA	Ferozsons Laboratories Limited
31.	31	PTCA GUIDE WIRE, HYDROPHILIC, NON TAPERED INTERMEDIATE SUPPORT, STRAIGHT TIP, POLYMER COATED	T.N.A 1. The specification i.e hydrophilic not mentioned in technical offer 2. Medical device registration certificate not attached 3. Punjab govt. experience not attached 4. Invalid Free sale certificate attached 5. Sample not provided for this product (in technical offer, they offered PTCA guide wires while the sample of PTCA Balloon provided) Rontis AG Lotus II 45 GILT 195/45XS Swiss	Bestech Enterprises
			T.A Choice Floppy Boston Scientific USA	Ferozsons Laboratories Limited
32.	32	PTCA GUIDE WIRE, HYDROPHILIC, NON TAPERED, INTERMEDIATE STRAIGHT TIP, MODERATE SUPPORT, 0.014" AND 175-195 CM	T.A BMW (.014", 190cm Hydrophilic coating, elastin distal core for kink resistance and durability, proximal stainless steel section for excellent push and torque transmission. Abbott vascular. USA/Europe	Digital Imaging Systems
			T.N.A 1. The specification i.e hydrophilic not mentioned in technical offer 2. Medical device registration certificate not attached 3. Punjab govt. experience not attached 4. Invalid Free sale certificate attached 5. Sample not provided Rontis AG Lotus II LT 195/60 NS	Bestech Enterprises

*[Handwritten signatures and marks at the bottom of the page]*

			Swiss	
33.	33	PTCA GUIDE WIRE, HYDROPHILIC, NON TAPERED, STIFF STRAIGHT TIP, MODERATE SUPPORT, 0.014" AND 175-195 CM POLYMER COATED ✓	T.A PT Graphic Boston Scientific USA	Ferozsons Laboratories Limited
34.	34	PTCA GUIDE WIRE, HYDROPHILIC, TAPERED, INTERMEDIATE SUPPORT STRAIGHT TIP POLYMER COATED ✓	T.A PT Graphic Boston Scientific USA	Ferozsons Laboratories Limited
35.	35	PTMC BALLOON WITH ACCESSORIES (INOUE TYPE) ✓	T.A Toray Japan	Verizon
36.	36	TEMPORARY PACEMAKER LEAD ✓	T.A Spike LC Fiab- SPA, Italy	A & E Medical
37.	37	TRANSFEMORAL SHEATH (SHEATH CATHETER WITH HAEMOSTATIC VALVE, GUIDE WIRE SPRING TYPE (J TYPE), DILATOR KINK RESISTANT, SELDINGER NEEDLE 18G WITH VALVE, IV CATHETER AND SCALPEL) ALL SIZES CE Approved Only ✓	T.A Accu Sheath Sung Won Medical, Korea	A & E Medical
			T.A Avanti+ Cordis USA	Global Marketing Services
			T.N.A 1. Filter leakage Femoral Sheath SCW Medicath, China	Health Tec
38.	38	TRANSRADIAL SHEATH (SHEATH CATHETER WITH HAEMOSTATIC VALVE, GUIDE WIRE (ST.TYPE), DILATOR KINK RESISTANT, SELDINGER NEEDLE 21G WITH VALVE, IV CATHETER AND SCALPEL)- ALL SIZES CE Approved Only ✓	T.A Accu Sheath Sung Won Medical, Korea	A & E Medical
			T.N.A 1. Sample not provided 2. Medical Device Registration certificate not attached 3. Free sale certificate not attached 4. Ambiguous authority letter attached (SHSMA & Lepu, Medical China both manufacturer authorized M/s Intek corporation for same products) Introducer Kit SHSMA Lepu, China	Intek Corporation
			T.A Avanti+ Cordis USA	Global Marketing Services.
			T.N.A 1. Filter leakage Radial Sheath SCW Medicath, China	Health Tec
39.	39	SNARE CATHETER ALL SIZES ✓	T.A EN Snare Catheter Merit Medical, USA	Cor-Med
40.	40	CD WRITEABLE / REWRITEABLE COMPATIBLE WITH MACHINE ✓	T.A Maxcell Taiwan	Ali Medical Store
			T.N.A 1. Sample not provided 2. Authority letter not attached 3. Punjab Govt. experience not attached Maxell Japan	Bilal International ✓
41.	41	DISPOSABLE LATEX GLOVES FOR EXAMINATION ALL SIZES (BOX OF 100) ✓	T.A Safehand Top Gloves SDN BHD, Malaysia	Ali Medical Store
			T.A Sempermed Semperit Group, Austria	Superior Healthcare International
			T.N.A 1. Authority letter from M/s A Feroz & Co to M/s Friends Enterprises attach while authority letter of M/s A Feroz & Co from the manufacturer not attached 2. Punjab govt. experience not attached 3. Free sale certificate not attached Safety	Friends Enterprises



			A Feroz & Co. Malaysia	
			T.N.A 1. The brand mentioned on authority letter was MYKU examination gloves while on technical offer the brand name was mentioned Primax Examination Gloves Primax Examination Gloves all Size Biomax, Malaysia	M Yousaf & Co.
			T.A Profeel PF WRP Asia Pacific Bhd. Malaysia	Ali Medical Store
			T.N.A 1. Ambiguous Authority letter (about validity) attached 2. Sufficient Punjab Govt. experience not attached 3. Sample not provided 4. Free sale certificate not attached Guarda Powder Free Rosner Mauby Meditrade GmbH Germany	Superior Healthcare International
42.	42	DISPOSABLE LATEX SURGICAL GLOVES STERILIZED (POWDER FREE) NO. 7	T.N.A 1. Sample not provided 2. Authority letter from M/s care & cure to M/s A.M Healthcare attached while from M/s care & cure to manufacturer (M/s Anhui Anyu Latex Products Co. Ltd. China) not attached While from M/s care & cure to another manufacturer name (M/s Yangzhou Goldenwell Medical device factory attached Cure Anhui Anyu Latex Products Co. Ltd. China	A.M Healthcare
			T.N.A 1. Agreement between M/s Searle Switzerland and M/s Searle Pakistan not attached while agreement between M/s Searle Pakistan and M/s TG Sdn BHD Malaysia attached While on physically the address mentioned on the gloves was M/s Searle Switzerland and M/s Searle Pakistan 2. Punjab govt. experience not attached 3. Free sale certificate not attached 4. Sample quality not up to mark Protiev Searle, Malaysia	Friends Enterprises
			T.A Profeel PF WRP Asia Pacific Bhd. Malaysia	Ali Medical Store
43	43	DISPOSABLE LATEX SURGICAL GLOVES STERILIZED (POWDER FREE) NO. 7.5	T.N.A 1. Ambiguous Authority letter (about validity) attached 2. Sufficient Punjab Govt. experience not attached 3. Poor quality (Rupture during handling) 4. Free sale certificate not attached Guarda Powder Free Rosner Mauby Meditrade GmbH Germany	Superior Healthcare International.
			T.N.A 1. Sample not provided 2. Authority letter from M/s care & cure to M/s A.M Healthcare attached while from M/s care & cure to manufacturer (M/s Anhui Anyu Latex Products Co. Ltd. China) not attached While from M/s care & cure to another manufacturer name (M/s Yangzhou	A M Healthcare

*[Handwritten signatures and scribbles at the bottom of the page]*



			Goldenwell Medical device factory attached Cure Anhui Anyu Latex Products Co. Ltd, China	
			T.N.A 1. Agreement between M/s Searle Switzerland and M/s Searle Pakistan not attached while agreement between M/s Searle Pakistan and M/s TG Sdn BHD Malaysia attached While on physically the address mentioned on the gloves was M/s Searle Switzerland and M/s Searle Pakistan 2. Sufficient Punjab govt. experience not attached 3. Free sale certificate not attached 4. Sample quality not up to mark Protix Searle, Malaysia	Friends Enterprises
44.	44	ECG CHEST ELECTRODE FOR ETT	T.N.A Poor conduction My Heart Qingdao Bright Medical, China	Ali Medical Store
			T.N.A Poor conduction Fiab Italy	Cardiac Care
			T.A Skintact F-60 Leonhard Lang Austria	Husnain Pharma
			T.N.A 1. Poor conduction 2. Punjab Govt. experience not attached 3. Country of origin on technical offer not mentioned Bio Protech INC	Flowtronix Systems
			T.N.A 1. Poor conduction 2. Punjab Govt. experience not attached 3. Authority letter not attached V Care China	Usman & Company
			T.N.A 1. Poor conduction 2. Discrepancy found in the name of the manufacture (The name of manufacturer mentioned on the authority letter was TYROLMED GmbH while on the free sale certificate issued by Austria (Country of origin) the manufacturer name was mentioned M/s Nessler Medizintechnik GmbH) Swaromed60mm Vermed Medical, Austria	Hakimsons Pvt Ltd
45.	45	ECG CHEST ELECTRODE FOR MONITORING - ADULT	T.N.A Poor conduction My Heart Qingdao Bright Medical Manufacturing CO. Ltd, China	Ali Medical Store
			T.N.A Poor conduction Fiab Italy	Cardiac Care
			T.A Skintact F-60 Leonhard Lang Austria	Husnain Pharma
			T.N.A 1. Poor conduction 2. Punjab Govt. experience not attached 3. Country of origin on technical offer not mentioned Bio Protech INC	Flowtronix Systems

Handwritten marks and signatures at the bottom of the page, including a large scribble and several initials.

			<p>T.N.A</p> <p>1. Poor conduction</p> <p>2. Discrepancy found in the name of the manufacture</p> <p>(The name of manufacturer mentioned on the authority letter was TYROLMED GmbH while on the free sale certificate issued by Austria (Country of origin) the manufacturer name was mentioned M/s Nessler Medizintechnik GmbH)</p> <p>Swaromed Adult</p> <p>Vermed Medical, Austria</p>	Hakimsons Pvt Ltd
			<p>T.N.A</p> <p>1. Poor conduction</p> <p>2. Punjab Govt. experience not attached</p> <p>3. Authority letter not attached</p> <p>V Care</p> <p>China</p>	Usman & Company
46.	46	ECG GEL LOCAL FOR ECG	<p>T.N.A</p> <p>Authority letter not attached</p> <p>Lizza White Gel</p> <p>Local, Pakistan</p>	Husnain Pharma.
			<p>T.N.A</p> <p>1. Authority letter not attached</p> <p>2. In technical offer brand mentioned (Aqua) while the sample of Aqua plus brand was provided</p> <p>Aqua</p> <p>Local, Pakistan</p>	Ibrahim Enterprises
			<p>T.A</p> <p>Aquasonic 100</p> <p>Parker Laboratories INC, USA</p>	Skardu Surgical Supplies
47.	47	ECG ROLL (3 CHANNEL) 63MM	<p>T.N.A</p> <p>Authority letter not attached</p> <p>Medichart</p> <p>China</p>	Usman & Company
			<p>T.A</p> <p>Telepaper 63mm</p> <p>Malaysia</p>	Friends Enterprises
			<p>T.A</p> <p>Lessa 63mm</p> <p>AB Medica Group S.A Spain</p>	Skardu Surgical Supplies.
			<p>T.N.A</p> <p>1. Sample not provided</p> <p>2. Authority letter not attached</p> <p>3. Punjab Govt. experience not attached</p> <p>Local</p> <p>Pakistan</p>	Bilal International.
48.	48	ECG ROLL (3 CHANNEL) 80MM	<p>T.A</p> <p>Lessa 80mm</p> <p>AB Medica Group S.A Spain</p>	Skardu Surgical Supplies.
			<p>T.N.A</p> <p>1. Sample not provided</p> <p>2. Authority letter not attached</p> <p>3. Punjab Govt. experience not attached</p> <p>Local</p> <p>Pakistan</p>	Bilal International
			<p>T.N.A</p> <p>Authority letter not attached</p> <p>Medichart</p> <p>China</p>	Usman & Company
			<p>T.A</p> <p>Telepaper 80mm</p> <p>Malaysia</p>	Friends Enterprises
49.	49	MANOMETER LINE-200CM LONG	<p>T.A</p> <p>Vygon Corp</p> <p>Europe/France</p>	Cardiac Care
			<p>T.N.A</p> <p>1. Manufacturer name and country of origin not mentioned on technical offer</p> <p>2. Punjab Govt. experience not attached</p> <p>3. Authority letter not attached</p> <p>4. Free sale certificate not attached</p> <p>5. Sample not according to specification</p>	Superior Healthcare International

~~XXXXXXXXXX~~      4      3      2

			provided Venous Flux	
			T.A Welcare ✓ Welford Manufacturing SDN BHD, Malaysia	Safeway System Pakistan
			T.A Pressure Line 200CM Bicakcilar Dis Ticaret A.S. Turkey	Ali Medical Store
50.	50	PRICKERS (BLOOD LENCET)	T.A Wellmed Sichuan Medicine and Health China	Husnain Pharma
			T.A Reco China	Ibrahim Enterprises
51.	51	STRIPS FOR GLUCOMETER (300 FOC GLUCOMETER REQUIRED COMPULSARY)	T.N.A 1. No experience with this institute 2. Validity of Authority letter not mentioned on authority letter from M/s I-Sens. Korea to M/s Rafi Sultan Caresens I-Sens. Korea	Friends Enterprises
			T.N.A 1. Invalid authority letter attached 2. Invalid free sale certificate attached On call Plus Acon USA	Popular International Pvt.Ltd
52.	52	SURGICAL TAPE (PAPER)-5 YARDS * 1" (4.5m)	T.N.A 1. Invalid authority letter from M/s Nichiban co.ltd Japan to M/s unifroze attached 2. Poor sticking Nichipore Surgical Tape Nichiban Co. Ltd. Japan	Ali Medical Store
			T.A Nitto Surgical tape Usman Enterprises, Pakistan	Husnain Pharma
			T.N.A 1. Punjab Govt. experience not attached 2. Authority letter not attached 3. Country of origin not mentioned on technical offer Homecare	Flowtronix Systems
53.	53	SURGICAL TAPE (PAPER)-5 YARDS * 3"	T.N.A 1. Sample not provided 1. Punjab Govt. experience not attached 2. Authority letter not attached 3. Country of origin not mentioned on technical offer Homecare	Flowtronix Systems
54.	54	TAPE FOR SKIN HYPO ALLERGIC ADHESIVE (TRANSPARENT) 3" X 10 YARDS (7.6cm x 9.1m)	T.N.A Authority letter not attached 3M	Flowtronix Systems
55.	55	THREE WAY STOP COCK - STANDARD	T.N.A 1. Not user friendly 2. Authority letter for this product not attached 3. Sufficient Punjab govt. experience not attached 4. Medical device registration certificate not attached Guangdong Baihe Medical Technology	Flowtronix Systems
			T.N.A 1. Sample not provided ✓ 2. Medical device registration certificate not attached Welcare Welford Manufacturing SDN BHD, Malaysia	Safeway System Pakistan

*[Handwritten signatures and scribbles at the bottom of the page]*

			<p><b>T.N.A</b></p> <p>1. Not user friendly</p> <p>2. Medical device registration certificate not attached</p> <p>3. Manufacturer name and country of origin not mentioned on Label</p> <p>4. Authority letter from M/s A Feroz &amp; Co to M/s Friends Enterprises attach while authority letter of M/s A Feroz &amp; Co from the manufacturer not attached</p> <p>5. Punjab govt. experience not attached</p> <p>Safety</p> <p>A Feroz &amp; Co, China</p>	Friends Enterprises
			<p><b>T.N.A</b></p> <p>Medical device registration certificate not attached</p> <p>3-Way Stop Cock</p> <p>Bicakilar Dis Ticaret A.S, Turkey</p>	Ali Medical Store
			<p><b>T.A</b></p> <p>Discofix</p> <p>B-Braun, Malaysia / Germany</p>	Husnain Pharma
			<p><b>T.N.A</b></p> <p>1. Punjab Govt. experience not attached</p> <p>2. Manufacturer name and country of origin not mentioned</p> <p>3. Authority letter not attached</p> <p>4. Medical device registration certificate not attached</p> <p>5. Free sale certificate not attached</p> <p>Venous Fixate-10</p>	Superior Healthcare International.
56.	56	URINE BAG 2000 ML	<p><b>T.N.A</b></p> <p>Sample not provided</p> <p>Icare ✓</p> <p>Welford Manufacturing SDN BHD, Malaysia</p>	Safeway System Pakistan
			<p><b>T.A</b></p> <p>MYKU Yrine Bag</p> <p>GC Medica, China</p>	M Yousaf & Co
			<p><b>T.A</b></p> <p>Foyomed 2000ml</p> <p>Foyomed Medical, China</p>	Hakimsons Pvt Ltd
			<p><b>T.N.A</b></p> <p>1. Brand name not mentioned</p> <p>2. Punjab Govt. experience not attached</p> <p>3. Authority letter not attached</p> <p>4. Free sale certificate not attached</p> <p>China</p>	Superior Healthcare International
			<p><b>T.A</b></p> <p>Wellmed</p> <p>Ningbo Greetmountain Medical</p> <p>China</p>	Husnain Pharma.
<b>List of PEADS</b>				
57.	57	ATRIAL SEPTOSOMY BALLOON	<p><b>T.N.A</b></p> <p>1. Medical Device Registration certificate not attached</p> <p>2. Free sale certificate not attached</p> <p>Septosomy Balloon</p> <p>Numed, USA</p>	Intek Corporation
58.	58	DELIVERY SYSTEM FOR STENTS (LONG SHEETH)	<p><b>T.N.A</b></p> <p>Medical device registration certificate not attached</p> <p>Cook Medical</p> <p>USA / Europe</p>	Verizon
59.	59	COARCTATION OF AORTA STENTS Curved (ALL SIZES)	<p><b>T.N.A</b></p> <p>1. Medical Device Registration certificate not attached</p> <p>2 Free sale certificate not attached</p> <p>CP Stent</p> <p>Numed, USA</p>	Intek Corporation

*[Handwritten signatures and initials at the bottom of the page]*

			T.A Bentley Germany	Verizon
60.	60 a.	DELIVERY SYSTEM FOR PDA DEVICES ALL SIZES (7FR, 8FR, 9 FR)	T.N.A 1. Ambiguous authority letter attached (SHSMA & Lepu. Medical China both manufacturer authorized M/s Intek corporation for same products) 2. No previous experience with this institute 3. Sample not provided 4. Free sale certificate not attached PDA Delivery System SHSMA Lepu, China	Intek Corporation
			T.A Delivery System for PDA AGA / Abbott Vascular USA / Europe	Digital Imaging Systems
			T.A Cocoon ✓ Vascular Innovation UK/Thailand	Easha Enterprises
			T.N.A 1. No previous experience with this institute Stecrease For PDA LifeTech, China	Cor-Med
61.	60 b.	PDA DEVICES ALL SIZES	T.N.A 1. Ambiguous authority letter attached (SHSMA & Lepu. Medical China both manufacturer authorized M/s Intek corporation for same products) 2. No previous experience with this institute 3. Sample not provided 4. Free sale certificate not attached Memopart PDA Occluder SHSMA Lepu, China	Intek Corporation
			T.A PDA amplatzer devices AGA / Abbott Vascular USA / Europe	Digital Imaging Systems
			T.A Cocoon ✓ Vascular Innovation UK/Thailand	Easha Enterprises
			T.N.A 1. No previous experience with this institute Cera PDA LifeTech, China	Cor-Med
62.	61 a.	DELIVERY SYSTEM FOR SEPTAL OCCLUDER DEVICES FOR ASD ALL SIZES	T.N.A 1. Ambiguous authority letter attached (SHSMA & Lepu. Medical China both manufacturer authorized M/s Intek corporation for same products) 2. No previous experience with this institute 3. Sample not provided 4. Free sale certificate not attached ASD Delivery System SHSMA Lepu, China	Intek Corporation
			T.A Delivery system for ASD AGA / Abbott Vascular USA / Europe	Digital Imaging Systems
			T.A Cocoon ✓ Vascular Innovation UK/Thailand	Easha Enterprises
			T.N.A 1. No previous experience with this institute Stecrease For ASD LifeTech, China	Cor-Med
63.	61 b.	ASD DEVICES (SEPTAL OCCLUDER) ALL SIZES	T.N.A 1. Ambiguous authority letter attached (SHSMA & Lepu. Medical China both manufacturer authorized M/s Intek corporation for same products)	Intek Corporation

*[Handwritten signatures and marks at the bottom of the page]*

			<p>2. No previous experience with this institute</p> <p>3. Sample not provided</p> <p>4. Free sale certificate not attached</p> <p>ASD Occluder</p> <p>SHSMA Lepu, China</p> <p>T.N.A</p> <p>Manufacturer name and country of origin was not mentioned on the technical offer</p> <p>Amplatzer ASD devices</p>	Digital Imaging Systems
			<p>T.A</p> <p>Cocoon</p> <p>Vascular Innovation UK/Thailand</p> <p>T.N.A</p> <p>1. No previous experience with this institute</p> <p>Cera ASD</p> <p>Life Tech, China</p>	Easha Enterprises
			<p>T.N.A</p> <p>1. Medical Device Registration not attached</p> <p>2. Free sale certificate not attached</p> <p>Radifocus Guide Wire</p> <p>Terumo, Japan</p>	Intek Corporation.
64.	62	Glide Wire 0.035 (Ordinary) Straight / curved Hydrophilic 150CM	<p>T.N.A</p> <p>1. Medical device registration certificate not attached</p> <p>2. Sample not provided</p> <p>Cook Medical</p> <p>USA / Europe</p>	Verizon
			<p>T.N.A</p> <p>No previous experience with this institute</p> <p>Poseidon Glide Wire 150cm</p> <p>SP Medical, Denmark</p>	Health Tec
			<p>T.N.A</p> <p>1. Medical Device Registration not attached</p> <p>2. Free sale certificate not attached</p> <p>Radifocus Guide Wire</p> <p>Terumo, Japan</p>	Intek Corporation.
65.	63	Glide Wire Exchange Length 0.035" & 0.025" 260CM Hydrophilic (Straight / Curved) 00.	<p>T.N.A</p> <p>No previous experience with this institute</p> <p>Poseidon Glide Wire 260cm</p> <p>SP Medical, Denmark</p>	Health Tec
			<p>T.N.A</p> <p>Medical device registration certificate not attached</p> <p>Cook Medical</p> <p>USA / Europe</p>	Verizon
			<p>T.N.A</p> <p>1. No previous experience with this institute</p> <p>Inquire Amplatzer</p> <p>Merit Medical, USA</p>	Cor-Med
66.	64	Guide Wire Extra Stiff Exchange Length 0.035" 260CM (Super Stiff)	<p>T.A</p> <p>Emerald</p> <p>Cordis USA</p>	Global Marketing Services
			<p>T.N.A</p> <p>Medical device registration certificate not attached</p> <p>Cook Medical</p> <p>USA / Europe</p>	Verizon
67.	65 a.	DELIVERY SYSTEM FOR VSD DEVICES ALL SIZES	<p>T.N.A</p> <p>1. Ambiguous authority letter attached (SHSMA &amp; Lepu, Medical China both manufacturer authorized M/s Intek corporation for same products)</p> <p>2. No previous experience with this institute</p>	Intek Corporation

			3. Sample not provided 4. Free sale certificate not attached VSD delivery system SHSMA Lepu, China	
			T.A Delivery system for VSD AGA / Abbott Vascular USA / Europe	Digital Imaging Systems
			T.N.A 1. No previous experience with this institute Cocoon Vascular Innovation UK/Thailand	Easha Enterprises
			T.A Steecease For MFO LifeTech, China	Cor-Med
68.	65 b.	VSD DEVICES ALL SIZES	T.N.A 1. Ambiguous authority letter attached (SHSMA & Lepu, Medical China both manufacturer authorized M/s Intek corporation for same products) 2. No previous experience with this institute 3. Sample not provided 4. Free sale certificate not attached VSD Occluder SHSMA Lepu, China	Intek Corporation
			T.A Amplatz VSD devices AGA / Abbott Vascular USA / Europe	Digital Imaging Systems
			T.N.A 1. No previous experience with this institute Cocoon Vascular Innovation UK/Thailand	Easha Enterprises
			T.A Konar MFO LifeTech, China	Cor-Med
69.	66	PULMONARY VALVULOPLASTY BALLOONS ALL SIZES (SOFT PROFILE)	T.N.A 1. Medical Device Registration not attached 2. Free sale certificate not attached Tyshak Numed, USA	Intek Corporation
			T.N.A 1. No previous experience with this institute 2. Sample not provided 3. Free sale certificate not attached Optitorque Terumo, Japan	Intek Corporation
70.	67	MP CATHETER ALL SIZES	T.A Infinity Cordis USA	Global Marketing Services
			T.N.A 1. Invalid drug registration certificate attached 2. Medical device registration certificate not attached Alvision Alvimedica, Turkey	ATCO Pharma International Pvt Ltd
			T.N.A 1. Stiff and hard material Pointer Pendraccare, Netherland	Health Tec
71.	68	NIH CATHETER ALL SIZES	T.A Infinity Cordis USA	Global Marketing Services
72.	69	PIGTAIL CATHETER ALL SIZES	T.N.A 1. No previous experience with this institute 2. Sample not provided 3. Free sale certificate not attached Optitorque Terumo, Japan	Intek Corporation


Handwritten signatures and initials at the bottom of the page, including a large signature on the left and several smaller initials or marks on the right.


			T.A Infinity Cordis USA	Global Marketing Services
			T.N.A 1. Invalid drug registration certificate attached 2. Medical device registration certificate not attached Alvison Alvimedica, Turkey	ATCO Pharma International Pvt Ltd
73.	70	EMBOLIZATION COILS ALL SIZES	T.N.A 1. Medical device registration not attached 2. No previous experience with this institute 3. Sample not provided Pushable Coils Boston Scientific USA	Ferozsons Laboratories Limited
			T.N.A Medical device registration certificate not attached Cook Medical USA / Europe	Verizon


The committee unanimously technically approved / technically not approved the brands of the above mentioned firms in each items.


Following items may be retendered.


Sr. No.	Award #	Name of Medicine
1.	2	MULLEN TRANSPETAL CATHETER INTRODUCER SHEATH
2.	3	BROKEN BROUGH TRANSPETAL NEEDLE
3.	8	COVERED STENT - ALL SIZES
4.	10	DISTAL PROTECTION DEVICE (FILTERTYPE) ALL SIZES
5.	12	FFR WIRE
6.	15	GUIDE WIRE EXCHANGE LENGTH 0.025" 260CM J TIP
7.	21	MICRO/SUPPORT HYBIRD CATHETER
8.	22	PERFUSION BALLOON - ALL SIZES
9.	31	PTCA GUIDE WIRE, HYDROPHILIC, NON TAPERED INTERMEDIATE SUPPORT, STRAIGHT TIP, POLYMER COATED
10.	51	STRIPS FOR GLUCOMETER (300 FOC GLUCOMETER REQUIRED COMPULSARY)
11.	53	SURGICAL TAPE (PAPER)-5 YARDS * 3"
12.	54	TAPE FOR SKIN HYPO ALLERGIC ADHESIVE (TRANSPARENT) 3" X 10 YARDS (7.6cm x 9.1m)
13.	57	ATRIAL SEPTOSOMY BALLOON
14.	58	DELIVERY SYSTEM FOR STENTS (LONG SHEETH)
15.	62	GLIDE WIRE 0.035 (ORDINARY) STRAIGHT / CURVED HYDROPHILIC 150CM
16.	63	GLIDE WIRE EXCHANGE LENGTH 0.035" & 0.025" 260CM HYDROPHILIC (STRAIGHT / CURVED)
17.	66	PULMONARY VALVULOPLASTY BALLOONS ALL SIZES (SOFT PROFILE)
18.	70	EMBOLIZATION COILS ALL SIZES


  
\_\_\_\_\_  
Electronic Equipment Care Taker,  
F.I.C. Faisalabad.

  
\_\_\_\_\_  
Pharmacist,  
F.I.C. Faisalabad.

  
\_\_\_\_\_  
AMS (Purchase/Stores),  
F.I.C. Faisalabad.

  
\_\_\_\_\_  
Dr. Abdul Razzaq Mughal,  
Assistant Professor of Cardiology  
(Paeds) FIC, Faisalabad

  
\_\_\_\_\_  
Dr. Rehan Riaz,  
Assistant Professor of Cardiology,  
FIC, Faisalabad.

  
\_\_\_\_\_  
Dr. Muhammad Hamid Saeed  
Head of Cardiology,  
FIC, Faisalabad.