

**MINUTES OF GRIEVANCE COMMITTEE MEETING
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD**

A meeting of Grievance Committee for product of CSSD Disposable Items for F.Y. 2020-21 was held on 14-10-2020 at 10:00AM in the Office of the Medical Superintendent, Faisalabad Institute of Cardiology, Faisalabad.

CSSD DISPOSABLE ITEMS

1. M/s Husnain Pharma

| Item # | Item Name | Brand Name | Manufactured By / Imported By | TAC Observations | GC Observation | Remarks |
|--------|---|------------------|-------------------------------|----------------------|-----------------|-------------------|
| 17 | High level instrument disinfectant solution (5 liter) | ReproDIS HLD4I | Medimark Scientific, UK | Sample not provided | Sample provided | Refer back to TEC |
| 18 | Hand Sanitizer Gel Form | Hi Clean Alcohol | Local Pakistan | *Sample not provided | Sample provided | Refer back to TEC |

2. M/s APSTA INT

| Item # | Item Name | Brand Name | Manufactured By / Imported By | TAC Observations | GC Observation | Remarks |
|--------|--|-------------------|---------------------------------|---|----------------------|----------|
| 13 | Alkaline Chemical for washer disinfectant (5liter) | Labomat MA Liquid | Laboratorium Dr. Deppe, Germany | *Impact not good *No shining after process | End user reservation | Rejected |
| 14 | Acid for washer disinfectant (5liter) | Labomat KS | Laboratorium Dr. Deppe, Germany | *Poor quality | End user reservation | Rejected |

Handwritten signatures in blue ink, including a large signature on the left and several smaller ones on the right.

3. M/s Mediland Pakistan Pvt. Ltd.

| Item # | Item Name | Brand Name | Manufactured By / Imported By | TAC Observations | GC Observation | Remarks |
|--------|---|----------------------------|-------------------------------|--|--|-------------------|
| 14 | Acid for washer disinfectant (5liter) | Gentinge Infection Control | UK | *Not compatible with machine because Alkaline not attached | End user reservation | Rejected |
| 15 | Lubricant for washer disinfectant (5 liter) | Gentinge Infection Control | UK | Sample not compatible | End user reservation | Rejected |
| 21 | Steam chemical indicator (class 6) | Gentinge Infection Control | USA | *Discolour after process | The firm provided sample for re-evaluation | Refer back to TEC |

4. M/s Cardiac Care

| Item # | Item Name | Brand Name | Manufactured By / Imported By | TAC Observations | GC Observation | Remarks |
|--------|---|-------------|-------------------------------|------------------------|----------------------|-------------------|
| 1 | Green Sterilization Wrapping paper (60 gms) 250 sheets/box (75x75) cm | E-Line | Italy | *Hard after processing | End user reservation | Rejected |
| 2 | Green Sterilization Wrapping paper (60 gms) 250 sheets/box (50x50) cm | E-Line | Italy | *Hard after processing | End user reservation | Rejected |
| 16 | Label Printing Gun | SPS Medical | USA | *Sample not provided | Sample provided | Refer back to TEC |



