

04/AMS/18
07/02/18

**MINUTES OF GRIEVANCE COMMITTEE MEETING
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD**

A meeting of Grievance Committee for Cardiac Surgical Disposables was held on 07.02.2017 (Wednesday) at 11:00AM in the Office of the Medical Superintendent, Faisalabad Institute of Cardiology, Faisalabad.

ANGIOGRAPHY & ANGIOPLASTY DISPOSABLES

1. M/s: Medica

Item #	Item Name	TAC Observations	Remarks
24	BLANKET ADULT COMPATIBLE WITH SORINS HYPOTHERMIA MACHINE	Govt. Experience not attached	Rejected

2. M/s Ali Medical Store

Item #	Item Name	TAC Observations	Remarks
6	DISPOSABLE FACE MASK	Middle layer of filter not breathing friendly Not porous filter	Rejected
27	CENTRAL VENOUS CATHERIZATION SET TRIPLE LUMEN 7FR	Guide wire long Indicator measures not clear Catheter long 20CM	Rejected
28	CENTRAL VENOUS CATHERIZATION SET SINGLE LUMEN 7FR	Small puncture needle & dilator	Rejected

3. M/s Iqbal & Company


Item #	Item Name	TAC Observations	Remarks
15	HEMODIALYSIS CATHETER DOUBLE LUMEN	Incomplete accessories	As the mentioned accessories are not demanded in the Tender documents so the decision referred back to TAC for Re-evaluation
17	HEMODIALYSIS KIDNEY FILTER	Authority letter not attached	Accepted Authority Letter Provided
26	CENTRAL VENOUS CATHERIZATION SET TRIPLE LUMEN 5.5FR	Incomplete accessories	As the mentioned accessories are not demanded in the Tender documents so the decision referred back to TAC for Re-evaluation
27	CENTRAL VENOUS CATHERIZATION SET TRIPLE LUMEN 7FR	Scalpel not present Catheter size long Slip through butterfly	As the mentioned accessories are not demanded in the Tender documents so the decision referred back to TAC for Re-evaluation





28	CENTRAL VENOUS CATHERIZATION SET SINGLE LUMEN 7FR	Scalpel not present Catheter size long Slip through butterfly CVP catheter not graduated	As the mentioned accessories are not demanded in the Tender documents so the decision referred back to TAC for Re-evaluation
----	---	---	--


4. M/s Flowtronix System


Item #	Item Name	TAC Observations	Remarks
13	STAB LINES APPROVED	Authority letter is only for aortic punch (IBC) attached	Rejected Invalid Authority Letter


Aysha Nasrullah
Pharmacist,
FIC, Faisalabad


Engr. Najam-ul-Hassan
Electronic Equipment Care Taker,
FIC, Faisalabad


Dr. Muhammad Sajid,
Deputy Medical Superintendent,
FIC, Faisalabad


Dr. Faisal Ali,
Consultant Cardiologist,
FIC, Faisalabad


Dr. Rafi Zulfiqar
Medical Superintendent,
FIC, Faisalabad