

11/AMS/18  
24/09/18

**MINUTES OF GRIEVANCE COMMITTEE MEETING  
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD**

A meeting of Grievance Committee for CSSD & Blood Bank Items was held on 22.09.2018 at 11:00AM in the Office of the Chairman Grievance Committee, Faisalabad Institute of Cardiology, Faisalabad.

**CSSD**

**1. M/s Total Technologies**

Item #	Item Name	Brand Quoted	TAC Observations	Firm Version/Grievance	GC Observation	Remarks
1	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (100x100) cm	Arvand Fan	T.N.A Not according to specification Hard after sterilization	The firm claims that the manufacturer of the quoted product is ISO, CE certified and being used in many hospitals throughout Pakistan.	As per discussion with end user, this wrapping paper has very rough texture that becomes hard after sterilization, which is liable to tear very easily. Moreover the firm fails to provide experience of any Govt. Institute.	Rejected
2	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (90x90) cm	Arvand Fan	T.N.A Not according to specification Hard after sterilization	-do-	-do-	Rejected
3	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (75x75) cm	Arvand Fan	T.N.A Not according to specification Hard after sterilization	-do-	-do-	Rejected
4	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (50x50) cm	Arvand Fan	T.N.A Not according to specification Hard after sterilization	-do-	-do-	Rejected
6	Flat Reel 7.5 cm x 200m (Packing Material)	Arvand Fan	T.N.A Film side shrink after process	-do-	As per discussion with end user, this film side shrink after process. Moreover the firm fails to provide experience of any Govt. Institute.	Rejected
7	Flat Reel 10cm x 200m (Packing Material)	Arvand Fan	T.N.A Film side shrink after process	-do-	-do-	Rejected
8	Flat Reel 15cm x 200m (Packing Material)	Arvand Fan	T.N.A Film side shrink after process	-do-	-do-	Rejected
9	Gussetted Reel 15cm x 100m (Packing Material)	Arvand Fan	T.N.A Film side shrink after process	-do-	-do-	Rejected
12	ETO Cartridge Gas (box of 12)	3M	T.N.A Sample not provided	-do-	-do-	Rejected

am h

*[Signature]*

## 2. M/s Quintex Medicals

Item #	Item Name	Brand Quoted	TAC Observations	Firm Version/Grievance	GC Observation	Remarks
1	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (100x100) cm	PMS	T.N.A Damaged, hard after sterilization	The firm claims that the manufacturer of the quoted product is ISO, CE and FDA certified and being used in many public and private hospitals throughout Pakistan including FIC, Faisalabad.	The end user not satisfied with the past experience of this product.	Rejected
2	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (90x90) cm	PMS	T.N.A Damaged, hard after sterilization	-do-	-do-	Rejected
3	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (75x75) cm	PMS	T.N.A Damaged, hard after sterilization	-do-	-do-	Rejected
4	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (50x50) cm	PMS	T.N.A Damaged, hard after sterilization	-do-	-do-	Rejected
5	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (30x30) cm	PMS	T.N.A Damaged, hard after sterilization	-do-	-do-	Rejected
6	Flat Reel 7.5 cm x 200m (Packing Material)	PMS	T.N.A Poor quality, hard after process (Film side), sealing strength not good	-do-	-do-	Rejected
7	Flat Reel 10cm x 200m (Packing Material)	PMS	T.N.A Poor quality, hard after process (Film side), sealing strength not good	-do-	-do-	Rejected
8	Flat Reel 15cm x 200m (Packing Material)	PMS	T.N.A Poor quality, hard after process (Film side), sealing strength not good	-do-	-do-	Rejected

W

M

G

M

P

9	Gussetted Reel 15cm x 100m (Packing Material)	PMS	T.N.A Poor quality, hard after process (Film side), sealing strength not good	-do-	-do-	Rejected
10	Gussetted Reel 40cm x 100m (Packing Material)	PMS	T.N.A Poor quality, hard after process (Film side), sealing strength not good	-do-	-do-	Rejected
11	Steam Indicator Tape (Lead free) (box of 48)	PMS	T.N.A Indicator not good after process	-do-	-do-	Rejected

## 3. M/s Three Stars

BLOOD BANK

Item #	Item Name	Brand Quoted	TAC Observations	Firm Version/Grievance	GC Observation	Remarks
2	Blood Lancets	China	T.N.A .Specific brand not mentioned	This product is being used in many hospitals all over the Punjab.	The technical bid shows that only country of origin is mentioned with no specific brand name.	Rejected
17	Torniquit All Brand	China	T.N.A .Specific brand not mentioned	-do-	-do-	Rejected
26	ADJUSTER 100-1000 ML	China	T.N.A .Specific brand not mentioned	-do-	do	Rejected
27	ADJUSTER 10-100 ML	China	T.N.A .Specific brand not mentioned	-do-	do	Rejected
28	ALCOHAL SWAB	China	T.N.A .Specific brand not mentioned	-do-	do	Rejected







## 4. M/s Akzan Enterprises

Item #	Item Name	Brand Quoted	TAC Observations	Firm Version/Grievance	GC Observation	Remarks
2	Blood Lancets	All Brand Pkt/200pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
3	Blue Tips Disposable	All Brand Pkt/500pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
6	Cover Slips 18x18mm All Brand Japan / Germany or equivalent	All Brand Pkt/200pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand perfect-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
7	Disposable Serum Cups	All Brand Pkt/500pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
8	Glass slide	All Brand Pkt/72pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Sail Brand-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
13	Plane test Tubes (Glass) 12X75mm	All Brand	T.N.A -Specific brand not mentioned	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
14	Plastishine All Brands	All Brand	T.N.A -Specific brand not mentioned	The firm has nominated brand Chinese Brand.	The technical offer can't be accepted at this level of bidding process.	Rejected
15	Sticking Bandage/sunny plast	All Brand Pkt/100pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Sunny Plast Brand	The technical offer can't be accepted at this level of bidding process.	Rejected
17	Torniquit All Brand	All Brand	T.N.A -Specific brand not mentioned	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
18	Yellow Tips Disposable All Brands	All Brand Pkt/1000pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
24	EDTA VACCUTAINER	China Pkt/100pcs	T.N.A -Sample not provided	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected


*[Handwritten signature]*

*[Handwritten signature]*

25	GEL & CLOT VACCUTAINER	China Pkt/100pcs	T.N.A Sample not provided	The firm has nominated brand Shangyu-China along with distributor letter	The technical offer can't be accepted at this level of bidding process.	Rejected
28	ALCOHAL SWAB	China Pkt/200pcs	T.N.A -Specific brand not mentioned	The firm has nominated brand Chinese Brand.	The technical offer can't be accepted at this level of bidding process.	Rejected



Mrs. Rabiah Karwal,  
Pharmacist,  
FIC, Faisalabad.



Mr. Zaman Ithikhar  
BME Engineer,  
FIC, Faisalabad.



Dr. Zulfiqar Haider,  
Associate Prof. of Cardiac Surgery,  
FIC, Faisalabad.



Dr. Mohsin Nazir,  
Prof. of Cardiology,  
FIC, Faisalabad.



Dr. Aiaz Akhtar Bhatti,  
Medical Superintendent,  
FIC, Faisalabad.