

Annexure A

SPECIAL TERMS & CONDITION FOR ABG'S MACHINE ON REAGENT BASIS

1. The offered machines should be CE / JIS / FDA approved with their detailed catalogue otherwise the bidder will straight away be rejected.
2. The bidder should provide national and international references of the offered machine.
3. The ABG's analyzer must have an inbuilt, non editable system of calculations of number of tests performed which can be printed.
4. Offered machine should have ability to perform following test / result / parameters.

➤ Lactate	Hb	B.P
➤ pH	So2	Patient Temperature
➤ PCO2	Cl	Bio data Entry
➤ PO2	Be. Ecf	
➤ Hct	Beb	
➤ Na	SBC	
➤ K		
➤ Ca		
➤ Hco3		

5. Successful bidder will be declared on cost of test & no of usefull parameters. Cost of test will be calculated on the basis of the number of tests performed i.e. per pack cell, reagent used by the machine for calibration/cleaning will not be included in total No. of tests.
6. The Successful bidder / contractor will provide five number of ABG's machines to FIC, Faisalabad. Next year contract will depend upon quality, supply and services feedback. The machine will remain the property of the successful bidder / contractor.
7. The successful bidder will also provide online UPS and batteries with each machine, Maintenance of UPS including batteries will also be the responsibility of the successful bidder. Institute will only provide power connection on site.
8. The bidder will be responsible to provide backup services within two hours. Incase of breakdown of system, if not corrected the problem within two hours then contractor will have to replace the machine.
9. The successful bidder will be responsible for every kind of spare parts, consumables or any other cleaning reagent involved for the functionality of the machines.
10. Incase of unsatisfactory backup services or poor performance of the machine, the institute will cancel the contract immediately and performance guarantee will be forfeited as penalty, coupled with blacklisting to the said contractor for doing further business with installation in future and same will be requested to Government of the Punjab.
11. Government tax will be approved as per Rules.
12. Any other instruction of the federal, provisional or local Government including boar issued on the subject at during currency of this agreement will be applicable.
13. The units of Ca+, Hb, Hco3 should be multiple i.e mg/dl, mol, mmol/l, g/mol/l etc.

